



Product Service

CERTIFICATE

No. Z1A 012906 0518 Rev. 00

Holder of Certificate: **DÜPERTHAL SICHERHEITSTECHNIK
GMBH & Co. KG**



DÜPERTHAL

Frankenstraße 3
63791 Karlstein
GERMANY

Certification Mark:



Product: **Safety cabinets**

Tested according to: DIN EN 14470-1:2004
EK5/AK4 21-01:2021
DIN EN 16121:2017
DIN EN 16122:2012
AfPS GS 2019:01 PAK

With this certificate, in its capacity as a notified GS body, TÜV SÜD Product Service GmbH issues the GS Mark to the named certificate holder for the declared product with the named model designation in accordance with section 20 ProdSG (German Product Safety Act). The product meets the safety and health requirements of section 20 (3) ProdSG. The manufacturer is reminded of their obligations in accordance with section 24 ProdSG. The certification marks shown above can be affixed on the product. It is not permitted to alter the certification marks in any way. In addition, the certificate holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier. All applicable requirements of the testing and certification regulations of TÜV SÜD Group have to be complied. For details see: www.tuvsud.com/ps-cert

Test report no.: 034-713251143-000

Valid until: 2027-08-16

Date, 2022-08-17

(Frank Feihle)

CERTIFICATE

No. Z1A 012906 0518 Rev. 00

Model(s): PREMIUM standard,
PREMIUM pro,
PREMIUM vario,
COMFORT standard

Parameters:

PREMIUM standard	
model type	approx. size without socket in mm (width x depth x height)
M	594 x 597 x 2037
ML	594 x 747 x 2037
XL	1194 x 597 x 2037

PREMIUM pro	
model type	approx. size without socket in mm (width x depth x height)
M	594 x 597 x 2037
ML	594 x 747 x 2037
XL	1194 x 597 x 2037

PREMIUM vario	
model type	approx. size without socket in mm (width x depth x height)
M	594 x 597 x 2037
XL	1194 x 597 x 2037

COMFORT standard	
model type	approx. size without socket in mm (width x depth x height)
M	594 x 597 x 2037
XL	1194 x 597 x 2037

Class of fire resistance: FWF 90

A detailed description of the specifications can be found in the test report.