



Product Service

CERTIFICATE

No. Z1A 012906 0527 Rev. 00

Holder of Certificate: **DÜPERTHAL SICHERHEITSTECHNIK
GMBH & Co. KG**



Frankenstraße 3
63791 Karlstein
GERMANY

Certification Mark:



Product: **Safety cabinets**

Tested according to: PPP 52125C:2023
DIN EN 16121:2025
DIN EN 16122:2012
AfPS GS 2019:01 PAK

With this certificate, in its capacity as a notified GS body, TÜV SÜD Product Service GmbH issues the GS Mark to the named certificate holder for the declared product with the named model designation in accordance with section 20 ProdSG (German Product Safety Act). The product meets the safety and health requirements of section 20 (3) ProdSG. The manufacturer is reminded of their obligations in accordance with section 24 ProdSG. The certification marks shown above can be affixed on the product. It is not permitted to alter the certification marks in any way. In addition, the certificate holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier. All applicable requirements of the Testing, Certification, Validation and Verification Regulations of TÜV SÜD Group have to be complied. For details see: www.tuvsud.com/ps-cert

Test report no.: 713370912

Valid until: 2030-08-05

Date, 2025-08-21

(Lisa Götz)

CERTIFICATE

No. Z1A 012906 0527 Rev. 00

Model(s): ACID C pro, ACID U pro, ACID B pro

Parameters:

ACID C pro	
model size	approx. size with socket in mm (width x depth x height)
S	600 x 609 x 1385
XS	1200 x 609 x 1385
M	600 x 609 x 2045
XL	1200 x 609 x 2045

ACID U pro	
model size	approx. size without socket in mm (width x depth x height)
S	600 x 609 x 596
L	1100 x 609 x 596

ACID B pro	
model size	approx. size without socket in mm (width x depth x height)
S	600 x 609 x 770
L	1100 x 609 x 770

Class of fire resistance: FWF 0

A detailed description of the specifications can be found in the test report.