



## CERTIFICATE

No. Z1A 012906 0523 Rev. 00

### Holder of Certificate:



#### DÜPERTHAL SICHERHEITSTECHNIK GMBH & Co. KG

Frankenstraße 3 63791 Karlstein GERMANY

**Certification Mark:** 



#### Product:

#### Safety cabinets

**Tested according to:** 

DIN EN 14470-2:2006 EK5/AK4 21-01:2021 DIN EN 16121:2024 DIN EN 16122:2012 AfPS GS 2019:01 PAK

With this certificate, in its capacity as a notified GS body, TÜV SÜD Product Service GmbH issues the GS Mark to the named certificate holder for the declared product with the named model designation in accordance with section 20 ProdSG (German Product Safety Act). The product meets the safety and health requirements of section 20 (3) ProdSG. The manufacturer is reminded of their obligations in accordance with section 24 ProdSG. The certification marks shown above can be affixed on the product. It is not permitted to alter the certification marks in any way. In addition, the certificate holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier. All applicable requirements of the Testing, Certification, Validation and Verification Regulations of TÜV SÜD Group have to be complied. For details see: www.tuvsud.com/ps-cert

Test report no.:

034-713310065-000

Valid until:

2029-07-02

Date, 2024-07-03

Horst Krísten

(Horst Kristen)

# CERTIFICATE

No. Z1A 012906 0523 Rev. 00

Model(s):

◆ CERTIFICAT

#### SUPREME standard, SUPREME pro

#### Parameters:

SUPREME standard	
Model type	approx. size in mm (width x depth x height)
S	595 x 610 x 1375
Μ	595 x 610 x 2030
L	895 x 610 x 2030
XL	1195 x 610 x 2030
XXL	1395 x 610 x 2030

SUPREME pro	
Model type	approx. size in mm (width x depth x height)
S	595 x 610 x 1375
М	595 x 610 x 2030
L	895 x 610 x 2030
XL	1195 x 610 x 2030
XXL	1395 x 610 x 2030

Class of fire resistance: FWF 90

A detailed description of the specifications can be found in the test report.