



Product Service

CERTIFICATE

No. Z1A 012906 0525 Rev. 00

Holder of Certificate: **DÜPERTHAL SICHERHEITSTECHNIK
GMBH & Co. KG**



Frankenstraße 3
63791 Karlstein
GERMANY

Certification Mark:



Product: **Safety cabinets**

Tested according to: DIN EN 14470-1:2023
EK5/AK4 21-01:2021
DIN EN 16121:2024
DIN EN 16122:2012
AfPS GS 2019:01 PAK

With this certificate, in its capacity as a notified GS body, TÜV SÜD Product Service GmbH issues the GS Mark to the named certificate holder for the declared product with the named model designation in accordance with section 20 ProdSG (German Product Safety Act). The product meets the safety and health requirements of section 20 (3) ProdSG. The manufacturer is reminded of their obligations in accordance with section 24 ProdSG. The certification marks shown above can be affixed on the product. It is not permitted to alter the certification marks in any way. In addition, the certificate holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier. All applicable requirements of the Testing, Certification, Validation and Verification Regulations of TÜV SÜD Group have to be complied. For details see: www.tuvsud.com/ps-cert

Test report no.: 713332584

Valid until: 2029-12-03

Date, 2024-12-04

Horst Kristen

(Horst Kristen)

CERTIFICATE

No. Z1A 012906 0525 Rev. 00

Model(s): COOL standard LL, COOL dual LL

Parameters:

COOL standard	
Model type	approx. size in mm (width x depth x height)
LL	894 x 791 x 2050

COOL dual	
Model type	approx. size in mm (width x depth x height)
LL	894 x 791 x 2050

Class of fire resistance: FWF 90

The in the safety storage cabinets integrated refrigerators and the exhaust air monitoring units were not part of the test.

A detailed description of the specifications can be found in the test report.